


Vergleichskörper 2 mit Zertifikat Magnetpulverprüfung

Art.-Nr. 134.002.100

Der Vergleichskörper 2 ist eine selbsthaltende Einheit, die keine externe Magnetisierung benötigt. Sie enthält zwei Stahlblöcke und zwei Permanentmagnete. Sie muss so kalibriert werden, dass die + 4-Markierung + 100 A/m und die - 4-Markierung - 100 A/m bedeuten.

Die Anzeigenlängen sind ein Maß für die Eignung. Die Anzeigen beginnen an den Enden und nehmen zur Mitte ab. Die bessere Eignung wird durch längere Anzeigen gekennzeichnet. Bewertet wird die Summe der Längen der linken und rechten Anzeigen.



Reference block 2 with certificate magnetic flux leakage methods

Art.-No. 134.002.100

Reference block type 2 is a self contained unit requiring no external magnetic field induction. It comprises 2 steel bars and two permanent magnets. It shall be calibrated such that the +4 mark represents +100 A/m and the -4 mark represents -100 A/m.

Indication lengths give a measure of performance. Indications start at the ends and decrease towards the centre. Increased length show better performance. Results shall be the cumulative length of the left and right hand indications